

I have a Coloplast Prosthetic Implant

To whom it may concern:

In case of emergency, please be aware that I have a Coloplast Penile Implant that **contains metal**. The implant is **MR conditional**.

See opposite side for details.

MR Conditional

- Static Magnetic field of 3 Tesla or less
- Spatial gradient field of 720 Gauss/cm or less
- Maximum whole-body averaged specification absorption rate (SAR) of 3W/kg for 15 minutes of scanning produced a temperature rise of less than 0.5°C.



For more information, please contact Coloplast
Surgical Support Department 800.258.3476



Coloplast

The coloplast logo is a registered trademark of Coloplast A/S.
©2019. All rights reserved. Coloplast Corp., Minneapolis, MN USA

PM-08544 08.19